

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer ODS Metering systems
Donk 6
2991 LE Barendrecht
The Netherlands

Measuring instrument An interruptible or non-interruptible **measuring system**
Manufacturer's mark or name : ODS Metering systems
Type : ODS Emerson Metering System
Accuracy class : 0.3; 0.5; 1.0 or 1.5
See §1.2 of the description
Intended for the measurement of : Oil and oil products, alcohol, chemicals, potable liquids, liquefied gasses under pressure and cryogenic liquids. See §1.2 of the description for the approved density range.

Further properties are described in the annex:

- Description T10235 revision 5.

Valid until 1 June 2030

Remarks

- The measuring system is approved for measuring mass, density and volume.
- This revision replaces the previous revisions.
- The documentation folder is obsolete.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
20 September 2021

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.