

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring systems (MID).

Manufacturer Toptech Europe N.V.
Nateus Business Park, Nieuwe weg 1 – Haven 1053
B-2070 Zwijndrecht
Belgium

In respect of An interruptible **measuring system** for liquids

Type : ML-Heliflu

Characteristics $Q_{min} - Q_{max}$: See paragraph 1.2
Minimum measured quantity : See paragraph 1.2
Accuracy class : 0,5
Environment class : M2 / E2
Temperature range liquid : -30 °C / +180 °C
Temperature range ambient : -25 °C / +55 °C
Intended for the measurement of : Hydrocarbons, alcohol, potable liquids and liquid chemicals

In the description number T10239 revision 0 further characteristics are described.

Valid until 30 November 2019

Description and documentation The measuring system is described in the description number T10239 revision 0 and documented in the documentation folder T10239-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122
NMI Certin, 30 November 2009



C. Oosterman
Head Certification Board