

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Emerson Process Management Flow B.V.
Neonstraat 1
6718 WX Ede
The Netherlands

Measuring instrument An interruptible or non-interruptible **measuring system** for liquids other than water
Manufacturer : Emerson
Type : MMI-MID 002
Accuracy class : 0.3; 0.5; 1.0; 1.5; 2.5
Intended for the measurement of : Oil and oil products, alcohol, potable liquids, chemicals, liquefied gasses under pressure and cryogenic liquids.

Further properties are described in the annexes:

- Description T10255 revision 10.
- Documentation folder T10255-3.

Valid until 10 June 2030

Remarks

- This revision replaces the previous revisions;
- The documentation folder is not changed;
- The measuring system is approved for measuring mass, density at metering conditions, density at base conditions, volume at metering conditions and volume at base conditions;
- The measuring system can be fixed or movable.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
30 November 2020

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.