| | | EU-type examination | |
|---|----------------------|--|--|
| | <u>N</u> MÌ | certificate | |
| Ŧ | | Number T10265 revision 18 Project number 3672016 Page 1 of 1 | |
| | Issued by | NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to: | |
| | Manufacturer | Emerson Process Management Flow B.V. Neonstraat 1 6718 WX Ede The Netherlands | |
| | Measuring instrument | A non-interruptible measuring system intended for delivery/ reception of marine fuel to/ from ships (barges and sea-going vessels). Manufacturer's mark or name : Emerson | |
| | | Type designation : MMI-MID-003 | |
| | | Accuracy class : 0.5 Destined for the measurement of : See § 1.2 of the description | |
| | | Further properties are described in the annexes: – Description T10265 revision 18. – Documentation folder T10265-16. – Annex 1: Bunker procedures for barges. – Annex 2: Bunker procedures for vessels. | |
| | Valid until | 15 June 2030 | |
| | Initially issued | 15 June 2010 | |
| | Remarks | The measuring system is approved for measuring mass, density at base conditions, volume at metering conditions and volume at base conditions. The measuring system can be fixed or movable. This revision replaces the previous revisions. The documentation folder replaces the previous documentation folder. | |
| | | | |

Issuing Authority

NMi Certin B.V. Thijseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl

NMi Certin B.V., Notified Body number 0122

16 May 2023

Certification Board

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <u>http://ec.europa.eu/growth/tools-</u> <u>databases/nando/</u> Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



ŧ