

EC-type examination Certificate

Number **T10272** revision 1 Project number 15200450 Page 1 of 1

Issued by Manufacturer	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to: Kemetyl Nederland B.V.				
	Industrieweg 30 3762 EK Soest The Netherlands				
Measuring instrument	An interruptible measuring instrument				
	Туре	+ +	Bulk filling machine		
	Destined for the measurement of Accuracy class Environment classes Temperature range liquid Temperature range ambient $Q_{min} - Q_{max}$ Minimum measured quantity	* * * * * * * * *	Coolants, windshield wash and other cleaning agents 0,5 M3 / E3 -10 °C / +50 °C -40 °C / +55 °C 50 - 1000 kg/min See § 1.2 of the descriptio	S + + + + + + + + + + + + + + + + + + +	
	Further properties are described in the annexes: – Description T10272 revision 1; – Documentation folder T10272-1.				
Valid until	12 May 2020				
+ Remarks + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + + +	This revision replaces the earlier ve folder.		$\begin{array}{cccccccccccccccccccccccccccccccccccc$	ation + + + + + + + + + + + + + + + + + + +	
+ Issuing Authority	Issuing Authority + + NMi certin B.V., Notified Body number 0122 + + + + + + + + + + + + + + +				
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	27 July 2015 C. Oosterman Head Certification Board				
NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl	This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability. The designation of NMi Certin BV.as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/	objecti within submis of NMi Reproc	concerned can lodge on against this decision, six weeks after the date of sion, to the general manager (see www.nmi.nl). Iuction of the complete ent only is permitted.	INSPECTION RVA 122	