


EC type-examination Certificate

Number **T10347** revision 0
Project number 11200052
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Shell Nederland Raffinaderij B.V. P.O. Box 3000 3190 GB Hoogvliet Rotterdam The Netherlands
Measuring instrument	An interruptible or non interruptible measuring system for liquids other than water Type : SNR/C-MID-001 Destined for the measurement of : Refined products Accuracy class : 0.3; 0.5; 1.0; 1.5; 2.5 Environment classes : Depending on the composition of the measuring instrument. Temperature range liquid : See § 1.2 of the description Temperature range ambient : See § 1.2 of the description $Q_{min} - Q_{max}$: See § 1.2 of the description Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10347 revision 0; – Documentation folder T10347-1.
Valid until	29 December 2021
Remarks	The measuring instrument is approved for measuring mass, density at metering conditions, density at base conditions, volume at metering conditions and volume at base conditions.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
29 December 2011



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.