

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	SGC Metering Gateway Business Park, Beancross Road FK3 8WX Grangemouth United Kingdom
Measuring instrument	An interruptible liquid <b>measuring system</b> Type : Oselvar oil metering package $Q_{min} - Q_{max}$ : 5892 – 11760 kg/h Accuracy class : 0.3 Environment classes : M3/E3 Temperature range liquid : -20 °C / +110 °C Temperature range ambient : -25 °C / + 55 °C Maximum pressure : 35 bar(g) Minimum measured quantity : 200 kg Intended for the measurement of : Oil and oil products  Further properties are described: – Description T10364 revision 0; – Documentation folder T10364-1.
Valid until	14 October 2021
Remarks	– The measuring system is only approved for measuring mass.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
14 October 2011

  
C. Oosterman  
Head Certification Board

**NMi Certin BV**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
PO Box 394  
3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document is permitted.