

EC type-examination certificate

Number **T10433** revision 0 Project number 11200816 Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:		
Manufacturer	FMC Kongsberg Metering Kirkegårdsveien 45 N-3616 Kongsberg Norway		
Measuring instrument	A non interruptible measuring		****
	Туре	: Liquid USM Meterin	g System
	Minimum – maximum flow rate	+ -: -300 - 8000 m ³ /h +	
	Minimum measured quantity	+ : 10 m^3 + + +	
	Maximum pressure	: 26,5 bar(g)	
	Accuracy class Environment classes	: 0,3 : M1 / E2	
	Temperature range liquid	: 0 °C – +70 °C	
	Temperature range ambient	: -25 °C − +30 °C	
	Destined for the measurement o		
	Further properties are described	in the annexes:	
	- Description T10433 revision 0;	* * * * * * * * * *	
	 Documentation folder T10433- 	1	
Valid until	11 February 2026		
+ Remarks + + + + +	The system is approved for meas	uring unconverted gross volu	ıme only.+ + + + +
	* * * * * * * * * * *		
Issuing Authority	NMi/Certin B.V., Notified Body	+ + + + + + + + + + + + + + + + + + +	
Issuing Authonity	11 February 2016		
	+ + + + + + + + + + + + + + + + + + + +		
	+++++++		
+ + + + + + + + 2	C. Oosterman Head Certification Board		
+ + + + + + + + + + + + + + + + + + +	This document is issued under the provision	Parties concerned can lodge	* * * * * * * *
Hugo de Grootplein 1	that no liability is accepted and that the	objection against this decision,	*
3314 EG Dordrecht The Netherlands	applicant shall indemnify third-party liability.	within six weeks after the date of submission, to the general manager	* (UN) * * *
+ T +31 78 6332332 + + + + + + certin@nmi.nl	The designation of NMi Certin B.V. as Notified Body can be verified at http://	of NMi (see <u>www.nmi.nl</u>). + + +	INSPECTION
www.nmi.nl	ec.europa.eu/enterprise/newapproach/nando/	Reproduction of the complete	
		document only is permitted.	+ + + + + + + + + + + + + + + + + + +