

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Emerson Process Management SAS 14 rue Edison Europarc du Chêne 69671 Bron Cedex France
Measuring instrument	A non-interruptible <b>measuring system</b> on a truck Type : MMI-MID-007 Destined for the measurement of : Cryogenic liquids Accuracy class : 2.5 Environment classes : M3 / E3 Temperature range liquid : See § 1.2 of the description Temperature range ambient : See § 1.2 of the description $Q_{min} - Q_{max}$ : See § 1.2 of the description Minimum measured quantity : See § 1.2 of the description  Further properties are described in the annexes: – Description T10476 revision 4; – Documentation folder T10476-2.
Valid until	13 April 2022
Remarks	The measurement system is only approved to measure mass.  This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
30 January 2015

  
C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete document only is permitted.