

EC type-examination Certificate

Number **T10494** revision 0 Project number 12200344 Page 1 of 1

Issued by	conformity modules mentioned in	Netherlands to perform tasks with respect to in article 9 of Directive 2004/22/EC, after suring instrument meets the applicable 22/EC, to:
Manufacturer	Van Cleemput Architecte Sprl Rue des Renards 11 6534 Thuin Belgium	
Authorised representative	Fabricom Kontichsesteenweg 25 2630 Aartselaar Belgium	
Measuring instrument	An interruptible measuring syst	stem for truck loading
• • • • • • • • • • •	Туре	: 91PO69 Defensie
	Destined for the measurement of	of + : Gasoil + + + + + + + + + + + + + + + + + + +
	Accuracy class Environment classes	: M3 / E2
	Temperature range liquid	: 0°C/+40°C
	Temperature range ambient	: -25 °C / +55 °C
	$Q_{min} - Q_{max}$: 250 - 1000 L/min
	Minimum measured quantity	: 500 L
	winning incusated quantity	* * * ****
	Further properties are described in – Description T10494 revision 0; – Documentation folder T10494-1	
+ + + + + + + + + + + + + + + + + + +	E November 2022	
Valid until	5 November 2023	
+ Issuing Authority +	NMi Certin B.V., Notified Body	y number 0122 + + + + + + + + + + +
	5 November 2013	
	C. Oosterman	
+ + + + + + + + + + + + + + + + + + + +	Head Certification Board	
NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl	This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability. The designation of NMi Certin BV.as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/	Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl). Reproduction of the complete document only is permitted
		document only is permitted.