

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	A. Silva Matos, SA Apartado 4 3740-909 Sever do Vouga Portugal
Measuring instrument	A non-interruptible measuring system Type : S10_X; S11_X Destined for the measurement of : See § 1.2 of the Description Accuracy class : 1,0 for LPG; 0,5 for other liquids Environment classes : M3 / E3 Temperature range liquid : -10 °C / +50 °C Temperature range ambient : -25 °C / +55 °C $Q_{min} - Q_{max}$: See § 1.2 of the Description Minimum measured quantity : See § 1.2 of the Description Further properties are described in the annexes: – Description T10499 revision 0; – Documentation folder T10499-1.
Valid until	11 October 2022

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
11 October 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.