

Issued by

NMi Certin B.V.
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Producer

Toptech Systems N.V.
Nieuwe weg 1
2070 Zwijndrecht
Belgium

Part

An interruptible **measuring instrument** for liquids other than water

Producer mark or name : Toptech Systems N.V.
Type designation : MultiLoad II/MM

Accuracy Class : 0,5; 1,0; depending on the applied measurement sensor.

Destined for the measurement of : Oil and oil products, alcohol, potable liquids, chemicals and liquefied gasses under pressure. See §1.2 of the description for the approved density range.

- Further properties and test results are described in the annex:
- Description T10505 revision 1.
 - Documentation folder T10505-2.

Valid until

30 April 2025

Remark

- This revision replaces the previous revision.
- The documentation folder replaces the previous documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
2 July 2021

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the producer shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.