

## EC type-examination certificate

Number **T10521** revision 1 Project number 11200816 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable

requirements of Directive 2004/22/EC, to:

Manufacturer \* \* \* FMC Kongsberg Metering

Kirkegårdsveien 45 N-3616 Kongsberg

Norway

Measuring instrument A non interruptible **measuring system** on a skid

Type : Knarr/Jordbær Metering System

 $Q_{min} - Q_{max}$  : 462 - 5081 m<sup>3</sup>/h

Minimum measured quantity : 10 m<sup>3</sup>
Maximum pressure : 19,6 bar(g)

Accuracy class : 0,3 Environment classes : M1 / E2

Temperature range liquid :  $-6 \,^{\circ}\text{C} - +70 \,^{\circ}\text{C}$ Temperature range ambient :  $-10 \,^{\circ}\text{C} - +30 \,^{\circ}\text{C}$ Destined for the measurement of : Crude Oil

Further properties are described in the annexes:

Description T10521 revision 1;Documentation folder T10521-1

Valid until 7 November 2024.

Remarks

This revision replaces the previous version, except for its documentation

folder.

Issuing Authority

NMi Certin B.V.

The Netherlands

Hugo de Grootplein 1

3314 EG Dordrecht

NMi/Certin B.V., Notified Body number 0122

11 February 2016

C. Øosterman

Head Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/ Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see <a href="https://www.nmi.nl">www.nmi.nl</a>).

Reproduction of the complete document only is permitted.

