

## EC-type examination Certificate

Number **T10688** revision 2 Project number SO16200796 Page 1 of 1

Issued by + + + + + NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable

requirements of Directive 2004/22/EC, to:

Manufacturer VIS Elektrotechniek B.V.

Adriaen Banckerstraat 10

3115 JE Schiedam The Netherlands

Measuring instrument An interruptible **measuring system** on a road tanker (truck)

Type : VISEL 3

Destined for the measurement of • • Oil and oil products, chemicals,

potable liquids with densities between 400 and 1400 kg/m<sup>3</sup>

Accuracy class : 0,5 Environment classes : M2 / E3

Temperature range liquid :  $-10 \,^{\circ}\text{C} / +50 \,^{\circ}\text{C}$ Temperature range ambient :  $-25 \,^{\circ}\text{C} / +55 \,^{\circ}\text{C}$ 

Q<sub>min</sub> - Q<sub>max</sub> : See the Description §1.2 Minimum measured quantity : See the Description §1.2

Further properties are described in the annexes:

Description T10688 revision 2;Documentation folder T10688-1

Valid until 13 February 2025

Remarks The measuring system is approved for measuring mass or volume.

This revision replaces the earlier versions, except for its documentation

folder.

14 March 2016

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl

www nmi nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

