

EU-type examination certificate

Number **T10813** revision 0
Project number: 16200434
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Honeywell Middle East Ltd. Abu Dhabi Cert Technology Park 45595 ABU DHABI United Arab Emirates
Measuring instrument	An interruptible measuring system for liquids other than water Type : Skid Type 1 – LPG Pipe Line Manufacturer's mark or name : Honeywell Destined for the measurement of : Under pressure liquefied propane, under pressure liquefied butane and mixtures thereof (LPG) Accuracy class : 1,0 Environment classes : M1 / E2 Temperature range liquid : +10 °C / +50 °C Temperature range ambient : -10 °C / +55 °C $Q_{min} - Q_{max}$: 667 – 2417 L/min Minimum measured quantity : 2000 L Maximum pressure (P_{max}) : 91 bar(g) Further properties are described in the annexes: – Description T10813 revision 0; – Documentation folder T10813-1.
Valid until	5 April 2027

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
5 April 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/>

Reproduction of the complete document only is permitted.