

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Talleres Angel Torres S.L.
P.I. Los Palillos, Calle 2, 27
Alcala De Guadaira
Sevilla
Spain

Measuring instrument An interruptible **measuring system** for liquids of low viscosity (< 20mPa.s)
other than water for static use or mounted on a road tanker.
Manufacturer's mark or name : Talleres Angel Torres
Type designation : TAT-S3xyzz, TAT-S4xyzz (see
paragraph 1.2 of the description for
the meaning of xyzz)
Accuracy class : 0.5
Destined for the measurement of : Liquids of low viscosity (< 20mPa.s)
other than water.

Further properties are described in the annexes:

- Description T11391 revision 0;
- Documentation folder T11391-1.

Valid until 25 February 2029

Remark The measuring system is approved for measuring volume.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
25 February 2019


C. Oosterman
Head Certification Board