

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer IvDam Process Control d.o.o.  
Koste Abraševića15  
11000 Belgrade  
Serbia

Measuring instrument An interruptible or non-interruptible **measuring system** for liquids other  
than water

Manufacturer's mark or name : IvDam

Type designation : IVD-FMSS; IVD-FMSM  
See chapter 1.2.11 for more details.

Accuracy class : 0.5 for other liquids; or  
1.0 for liquefied gases under  
pressure (LPG).

Destined for the measurement of : Oil and oil products, alcohol,  
potable liquids, chemicals in liquid  
state, or liquefied gases under  
pressure (LPG).

Further properties are described in the annexes:  
– Description T12337 revision 0.  
– Documentation folder T12337-1.

Valid until 21 April 2033

Initially issued 21 April 2023

Remarks

- The measuring system is approved for measuring mass, density and volume at actual conditions, standard volume and standard density when using the Micro Motion Coriolis meters.
- The measuring system is approved for volume at actual conditions and standard volume when using the Brodie or FMC PD-meters.

Issuing Authority

**NMI Certin B.V., Notified Body number 0122**  
21 April 2023

Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the manufacturer  
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [https://  
ec.europa.eu/growth/tools-databases/nando/](https://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally signed  
and sealed. The digital signature  
can be verified in the blue ribbon  
at the top of the electronic version  
of this certificate.