

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Endress+Hauser B.V.
Nikkelstraat 6
1411 AJ Naarden
The Netherlands

Measuring instrument An interruptible or non-interruptible **measuring system** for liquids other
than water, based on a dynamic mass (Coriolis) meter.
Manufacturer's mark or name : Endress + Hauser
Type designation : MID 001
Accuracy class : 0,3; 0,5; 1,0 or 1,5
Destined for the measurement of : Oil and oil products, chemicals,
potable liquids and liquefied gasses
under pressure with densities
between 400 and 1400 kg/m³.

Further properties are described in the annexes:

- Description T10221 revision 10;
- Documentation folder T10221-3.

Valid until 15 December 2029

Initially issued 15 December 2009

Remarks

- This revision replaces the previous versions, except for its documentation folder.
- The measuring system is approved for measuring mass, density and volume.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
8 May 2026

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the manufacturer
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at
<https://webgate.ec.europa.eu/single-market-compliance-space/home>

Reproduction of the complete
document only is permitted.

This document is digitally
signed and sealed. The digital
signature can be verified in the
blue ribbon at the top of the
electronic version of this
certificate.