

EU-type examination Certificate

Number **T10235** revision 7
Project number 3978650
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

ODS Metering Systems
Nikkelstraat 7
2984 AM Ridderkerk
The Netherlands

Measuring instrument

An interruptible and non-interruptible **measuring system** for liquids other
than water

Manufacturer's mark or name : ODS Metering Systems

Type designation : ODS Emerson Metering System and
ODS Coriolis Metering System

Accuracy class : 0.3; 0.5; 1.0 or 1.5
See 1.2 of the description

Destined for the measurement of : Oil and oil products, alcohol,
chemicals, potable liquids,
liquefied gases under pressure and
cryogenic liquids.
See chapter 1.2 of the description
for the approved density range.

Further properties are described in the annex:
– Description T10235 revision 7.

Valid until

1 June 2030

Initially issued

1 June 2010

Remarks

- This revision replaces the previous version(s).
- The documentation folder is obsolete.
- The measuring system is approved for measuring mass, density and
volume.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
8 July 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.