

EU-type examination Certificate



Number T10972 revision 1 Project number 3849692 Page 1 of 1

Issued by NMi Certin B.V.,

> designated and notified by the Netherlands to perform tasks with respect to conformity asssessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets

the applicable requirements of Directive 2014/32/EU, to:

Manufacturer **Turbines Incorporated**

Highway 283 North Altus, OK 73521

United States of America

Measuring instrument A non-interruptible or interruptible **measuring system** for liquids other

than water

Manufacturer's mark or name Turbine Inc.

Type designation Mobile or stationary liquid

measuring system

Accuracy class 1.5; 2.5

Destined for the measurement of Cryogenic liquids or liquefied gases

under pressure at a temperature

below -10 °C

Further properties are described in:

Description T10972 revision 1;

Documentation folder T10972-1.

Valid until 19 October 2028

Initially issued 19 October 2018

Remarks - This revision replaces the previous version(s), except for its documentation

folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122 26 August 2024

Certification Board

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party

The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T+31 88 636 2332 certin@nmi.nl www.nmi.nl

liability.