

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Endress Hauser Elektronik Sanayi ve Ticaret A.S.
Cevizli Mah. Tansel Cad. No:76, Civil Kule, Kat:3
34846 Istanbul
Türkiye

Measuring instrument An interruptible or non- interruptible **measuring system** for liquids other
than water, based on a dynamic mass (Coriolis) meter.
Manufacturer's mark or name : Endress + Hauser
Type designation : TR FMS 01
Accuracy class : 0,3; 0,5; 1,0; 1,5
Destined for the measurement of : Oil and oil products, chemicals,
potable liquids and liquefied gasses
under pressure with densities
between 400 and 1400 kg/m3.

Further properties are described in:

- Description T12836 revision 0;
- Documentation folder T12836-1.

Valid until 15 May 2034

Initially issued 15 May 2024

Remarks - The measuring system is approved for measuring mass, density and
volume at actual conditions, standard volume and standard density.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
8 August 2024

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.