

## EC-type examination certificate

Number **T10164** revision 2 Project number SO15203100 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable

requirements of Directive 2004/22/EC, to:

Manufacturer + + + Thermo Fisher Scientific

501 90<sup>th</sup> Ave NW

Minneapolis, MN 55433 United States of America

In respect of A model of a Continuous totalizing automatic weighing instrument,

intended to be used as a single speed belt weigher.

Type : Microtech 2200 series

Characteristics Accuracy class (OIML R50) : 2, 1 and 0,5

 $Q_{max}$  :  $\leq$  25000 t/h Mechanical environment class : not applicable

Electromagnetic environment class : E2

Temperature range : -10 °C / +40 °C, 85% humidity at 40 °C

non-condensing

In the description number T10164 revision 2 further characteristics are

described.

Valid until + + + + 17 September 2018

Description and The instrument is

Description and The instrument is described in the description number T10164 revision 2 documentation and documented in the documentation folder T10164-2, appertaining to this

EC-type examination certificate.

Remark This revision replaces the earlier versions, including its documentation folder.

ssuing Authority NMi Certin B.V., Notified Body number 0122

4 December 2015

. Oosterman

Head Certification Board

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

