

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Technipes S.r.l.  
Via Del Gelso, 12  
47822 Santarcangelo di Romagna  
Italy

In respect of A model of an **Automatic Gravimetric Filling instrument** intended to be used as cumulative weigher, subtractive weigher or single load weigher.  
Type: TE/ -series

Characteristics Reference accuracy class X (OIML R61) Ref(1)  
Mechanical environment class not applicable  
Electromagnetic environment class E2  
Temperature range +5 °C / +40 °C, 85% humidity at 40 °C non-condensing

In the description number T10223 revision 2 further characteristics are described.

Valid until 1 July 2019

Description and documentation The instrument is described in the description number T10223 revision 2 and documented in the documentation folder T10223-2, appertaining to this EC-type examination certificate.

Remarks This revision replaces the earlier versions, including its documentation folder.

The Notified Body no. 0122  
NMI Certin, 26 May 2010

  
C. Oosterman  
Head Certification Board