

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).
Manufacturer	Rocar-Tech Garage Equipment B.V. Granaatstraat 55 7554 TN Hengelo The Netherlands
In respect of	A model of an <b>Exhaust Gas Analyzer</b> . Type : 810
Characteristics	Accuracy class 0 Electromagnetic environment class E1 Temperature range +5 °C / +40 °C  In the description number T10468 revision 0 further characteristics are described.
Valid until	3 May 2022
Description and documentation	The instrument is described in the description number T10468 revision 0 and documented in the documentation folder T10468-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122  
NMi Certin, 3 May 2012

C. Oosterman  
Head Certification Board