

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant Mettler-Toledo  
1150 Dearborn Drive  
Worthington, Ohio 43085  
USA

In respect of A class **(III)**, electronic, single -interval **non-automatic weighing instrument**.  
Manufacturer : Mettler Toledo  
Type : SW ... Series

Characteristics  $n \leq 3000$  divisions  
 $3 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$  or  $6 \text{ lb} \leq \text{Max} \leq 300 \text{ lb}$   
 $e \geq 1 \text{ g}$  or  $e \geq 0.002 \text{ lb}$

In the description number T5529 revision 2 further characteristics are described.

Valid until 10 June 2019

Description and documentation The instrument is described in the description number T5529 revision 2 and documented in the documentation folder T5529-2, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier version, except for its documentation folder.

The Notified Body No. 0122  
NMI Certin, 10 June 2009

  
C. Oosterman  
Head Certification Board

**NMI Certin BV**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
PO Box 394  
3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.