

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant Avery Berkel
Foundry Lane
Smethwick
West Midlands
B67 9DF
United Kingdom

In respect of A class **(III)**, electronic, single- or multi-interval **non-automatic weighing instrument**, intended to be used for direct sales to the public.
Manufacturer : Avery Berkel, Avery, Berkel, Weighmaster Brecknell or Salter Brecknell
Type : FX, MC, B, G, BK, HL, S-series or SL-1200, SL-47xx

Characteristics $n \leq 7500$ divisions (single-interval instrument)
 $n \leq 3000$ divisions (per partial weighing range)
 $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$,
 $e \geq 1 \text{ g}$
maximum of two partial weighing ranges

In the description number T5532 revision 11 further characteristics are described.

Valid until 25 May 2019

Description and documentation The instrument is described in the description number T5532 revision 11 and documented in the documentation folder T5532-5, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier version, except for its documentation folder.

The Notified Body No. 0122
NMI Certin, 25 May 2009

C. Oosterman
Head Certification Board