

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 90/384/EEC on non-automatic weighing instruments.
Manufacturer	Acom Inc. 679-1, Yugyo-ri, Gunnae-myun, Pocheon-si, Kyunggi-do, 487-872, Korea
In respect of	A class (II) , electronic, single interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer : ACOM Type : PC-100
Characteristics	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 2 \text{ g}$ $n \leq 3000$ divisions Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ In the description number T5657 revision 2 further characteristics are described.
Valid until	18 January 2020
Description and documentation	The instrument is described in the description number T5657 revision 2 and documented in the documentation folder T5657-2, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier versions except for its documentation folder.

The Notified Body Number 0122
NMI Certin B.V., 18 January 2010



C. Oosterman
Head Certification Board