

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Mettler-Toledo AG Im Langacher 8606 Greifensee Switzerland
In respect of	A class <b>I</b> electronic, single-interval, <b>non-automatic weighing instrument</b> . Manufacturer mark/name: Mettler-Toledo Type : AX..., MX..., UMX...
Characteristics	Max $\leq$ 510 g e = 1 mg d $\leq$ e $\leq$ 10000 d n $\leq$ 510000 divisions Temperature range +10 °C / +30 °C In the description number T5708 revision 2 further characteristics are described.
Valid until	27 October 2020
Description and documentation	The instrument is described in the description number T5708 revision 2 and documented in the documentation folder T5708-1, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier versions, except for its documentation folder.

The Notified Body no. 0122  
NMI Certin, 27 October 2010

  
C. Oosterman  
Head Certification Board