

## EC type-approval certificate

Number T5908 revision 4 Project number 11200373 Page 1 of 4

NMi Certin B.V. Issued by

> Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC

with

Manufacturer Teraoka Seiko Co. Ltd.

13 12 Kugahara, 5 Chome Ohta ku, Tokyo 146

A class  $\stackrel{
m (III)}{
m (III)}$ , electronic, single- or multi-interval, **non-automatic weighin**g In respect of

instrument, intended to be used for direct sales to the public.

Teraoka Manufacturer mark/name: DS-788.. Type

 $3kg \le Max \le 30kg$ Characteristics

e ≥ 1g

n ≤ 3000 divisions (per partial weighing range)

Maximum of 2 partial weighing ranges Temperature range -10 °C / +40 °C

In the description number T5908 revision 4 further characteristics are described

Valid until 3 July 2021

Description and The instrument is described in the description number T5908 revision 4 and documentation documented in the documentation folder T5908-4, appertaining to this

EC type-approval certificate.

This revision EC type-approval certificate replaces the earlier Remarks

The Notified Body no. 0122 NMi Certin, 3 July 2011

Oosterman

**Head Certification Board** 

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability

The designation of NMi Certin BV. as Notified Body can be verified at http://

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete

