

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

By virtue of Article 11a, second part, of the "IJKwet" (Dutch Weights and Measures Law).

Applicant Arkema Vlissingen B.V. (formerly ATOFINA Vlissingen B.V.)  
Europaweg Zuid 2  
4389 PD VLISSINGEN  
The Netherlands

Submitted **A measuring system for liquids other than water** (based on a Coriolismeter)

Manufacturer : ATOFINA  
Designation : AVI

Characteristics Qmin – Qmax : see Description point 1.2  
Intended for the measurement of : chemical products in liquid state  
Environmental class : C  
Accuracy class : 0,5

In the Description T5926 Revision 2 the additional characteristics are given.

Description and Documentation The measuring system is described in de Description no. T5926 Revision 2; the documentation folder T5926–1 belonging to this Appendix to Type Approval remains unchanged.

Dordrecht, 15 December 2011.  
NMI Certin B.V.



C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
P.O. Box 394, 3300 AJ Dordrecht, NL  
phone +31 78 6332 332  
fax +31 78 6332 309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI B.V. (see "Regulation objection and appeal against decisions of NMI B.V.")

NMI B.V., chamber of comm. no. 27.228.701  
NMI Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

Reproduction of the complete document is permitted. Parts of this document may only be reproduced after written permission.