

## EC type-approval certificate

Number **T5955** revision 10 Project number 11219PAP Page 1 of 5

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC on non-automatic weighing instruments.

with

Manufacturer Ishida Co. Ltd.

44 Sanno-cho, Shogoin,

Sakyo-ku, Kyoto-city, 606-8392

Japan

In respect of A class (III),

A class (III), electronic, single- or multi-interval, non-automatic weighing

instrument, intended to be used for direct sales to the public.

Manufacturer mark/name: Ishida Co. Ltd.

Type : ASTRA-XT or AC-4000...

Characteristics

Single-interval	Multi-interval
6 kg ≤ Max ≤ 15 kg	6 kg ≤ Max ≤ 15 kg
n ≤ 6000 divisions	$n \le 3000$ divisions per partial weighing range,
+ + + + + + + + + + + +	<ul> <li>maximum of two partial weighing ranges</li> </ul>
$+ + + + + e \ge 2g + + + + +$	$+ + + + + + + e \ge 2g + + + + + + +$
+5 °C / +35 °C	-5 °C / +40 °C
T ≤ 84% of Max	$T < Max_1$

In the description number T5955 revision 10 further characteristics are described.

Valid until 17 October 2021

Description and The instrument is described in the description number T5955 revision 10 and documentation documented in the documentation folder T5955-6, appertaining to this

EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, except for its

documentation folder.

The Notified Body no. 0122 NMi Certin, 47 October 2011

C, Ooste man

**Head Certification Board** 

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

