

EC type-approval certificate

Number **T6027** revision 3 Project number SO12200678 Page 1 of 4

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands	
The Council Directive 2009/23/EC on non-automatic weighing instruments.	
CAS Corporation #19, Ganap-ri, Gwangjuk-Myoun Yangju-gun, Geonggi-do Republic of South Korea	
A class III, electronic, multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer mark/name: CAS Corporation Type : AP series	
$ \begin{array}{l} 6 \ kg \leq Max \leq 30 \ kg \\ e \geq 1 \ g \\ n \leq 3000 \ divisions \ (per \ partial \ weighing \ range) \\ Maximum \ of \ two \ partial \ weighing \ ranges \\ Temperature \ range \ -10 \ ^{\circ}C \ / \ +40 \ ^{\circ}C \\ In \ the \ description \ number \ T6027 \ revision \ 3 \ further \ characteristics \ are \ described. \end{array} $	
This revision EC type-approval certificate replaces the earlier versions, except for its + + documentation folder.	
April 2012	
This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability. The designation of NMi Certin BV. as Notified Body can be verified at http:// ec.europa.eu/enterprise/newapproach/nando/ Reproduction of the complete document only is permitted	
	Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands The Council Directive 2009/23/EC on non-automatic weighing instruments. CAS Corporation #19, Ganap-ri, Gwangjuk-Myoun Yangju-gun, Geonggi-do Republic of South Korea A class (), electronic, multi-interval, non-automatic weighing instrument, intended to be used for direct sales to the public. Manufacturer mark/name: CAS Corporation Type : AP series 6 kg ≤ Max ≤ 30 kg e ≥ 1 g n ≤ 3000 divisions (per partial weighing range) Maximum of two partial weighing ranges Temperature range - 10 °C / +40 °C In the description number T6027 revision 3 further characteristics are described. 16 April 2022 The instrument is described in the description number T6027 revision 3 and documented in the documentation folder T6027-3, appertaining to this EC type-approval certificate to unentation folder. rdy no. 0122 April 2012 on Board