

## EC type-approval certificate

Number **T6076** revision 7 Project number SO12200600 Page 1 of 4

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC on non-automatic weighing instruments.

with

Manufacturer Ishida Co. Ltd.

44, Sanno-cho, Shogoin

Sakyo-ku, Kyoto-city, 606-8392

Japan

In respect of A class (III), electronic, single-or multi-interval non-automatic weighing

instrument, intended to be used for direct sales to the public or not to be used for

direct sales to the public.

Manufacturer mark/name: Ishida or Cobos Precision S.L.
Type : IPC, IPC-WP, JC or JC-WP

Characteristics 1.5 kg  $\leq$  Max  $\leq$  30 kg

e ≥ 1 g

 $n \le 3000$  divisions for single interval instruments

 $n \le 1500$  divisions (per partial weighing range) for multi interval instruments

maximum of two partial weighing ranges

 $T \le - Max_1$ 

Temperature range -5 °C / +40 °C

In the description number T6076 revision 7 further characteristics are described.

Valid until 9 July 2022

Description and The instrument is described in the description number T6076 revision 7 and documentation documented in the documentation folder T6076-5, appertaining to this

EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, except for its

documentation folder.

The Notified Body no. 0122 NMi Certin, 9 July 2012

C. Oosterman Head Certification Boa

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

