

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Dräger Medical Systems Inc. 3135 Quarry Road Telford, PA 18969 United States of America
In respect of	A class (III) , electronic, single-interval, non-automatic weighing instrument , intended to be used for medical applications. Manufacturer mark/name: Dräger Medical Infant Care or Dräger Medical Systems Type : C2000 or Isolette 8000
Characteristics	Max \leq 7 kg e \geq 10 g n \leq 700 divisions Temperature range 10°C / 40°C In the description number T6621 revision 1 further characteristics are described.
Valid until	8 October 2014
Description and documentation	The instrument is described in the description number T6621 revision 1 and documented in the documentation folder T6621-1, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier version, except for its documentation folder.

The Notified Body no. 0122
NMi Certin, 27 October 2011



C. Oosterman
Head Certification Board