

Issued by NMI Certin B.V.  
 Hugo de Grootplein 1  
 3314 EG Dordrecht  
 The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer Descom Scales Mfg. Co. Ltd.  
 4-12 Wonmi-dong, Wonmi-ku  
 Buchon-city, Kyungki-do  
 Korea

In respect of A class **I** or **II**, electronic, self-indicating **non-automatic weighing instrument**.

Manufacturer : Descom  
 Brand name : Vibra  
 Type : AJ[H], AJ[H]P, AJ for carat (CT), SJ or SJP

Characteristics

Class <b>I</b>	Class <b>II</b>
$n \leq 82000$ divisions	$n \leq 42000$ divisions
$500 \text{ g} \leq \text{Max} \leq 6200 \text{ g}$ or $2500 \text{ ct} \leq \text{Max} \leq 31000 \text{ ct}$	$50 \text{ g} \leq \text{Max} \leq 12 \text{ kg}$ or $250 \text{ ct} \leq \text{Max} \leq 60000 \text{ ct}$
$e \geq 0,01 \text{ g}$ or $e \geq 0,1 \text{ ct}$	$e \geq 0,01 \text{ g}$ or $e \geq 0,1 \text{ ct}$
$e = d$ or $e = 10d$	$e = d$ or $e = 10d$

Temperature range +10 °C / +30 °C  
 In the description number T7086 revision 1 further characteristics are described.

Valid until 8 December 2016

Description and documentation The instrument is described in the description number T7086 revision 1 and documented in the documentation folder T7086-2, appertaining to this EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier version, including its documentation folder.

The Notified Body no. 0122  
 NMI Certin, 17 June 2011

  
 C. Oosterman  
 Head Certification Board