

Issued by NMI Certin B.V.
 Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer A.S de Ridder beheer Haaften B.V.
 Molenstraat 21
 4175 CA Haaften
 The Netherlands

In respect of An electronic, single-interval, single-range **non-automatic weighing instrument**.
 Manufacturer : De Ridder Weegschalen & Kasregisters B.V.
 Riba Weegtechniek B.V.
 Type : HJ(R)-..K.. and HJ(R)-CE series
 "R" = model with internal calibration mass

Characteristics

Type	HJ(R)-..K..	
Accuracy class	(II)	
Max	17 ~ 62 kg	85000 ~ 310000 ct
e ≥	1 g	5 ct
d ≥	d = e or d = 0,1 e	
n	≤ 62000	
Tare	≤ 100% of Max	
Temperature range	+5 °C / +35 °C	

Type	HJ(R)-CE & LN..(R)..		
Accuracy class	(I)	(II)	
Max	6200 g	1,2 kg ~ 15 kg	6000 ~ 75000 ct
e ≥	0,1 g	0,1 g	1 ct
d ≥	d = e or d = 0.1 e		
n	≤ 62000	≤ 42000	≤ 31000
Tare	≤ 100% of Max		
Temperature range	+10 °C / +30 °C		

EC type-approval certificate

Number **T7345** revision 1
 Project number 11200151
 Page 2 of 7

Type	HJ(R)-CE & LN..(R)..		
Accuracy class	(I)	(II)	
Max	620 g	220 g ~ 420 g	1100 ~ 3100 ct
e ≥	0,01 g		0,1 ct
d ≥	d = e or d = 0,1 e		
n	≤ 62000	≤ 42000	≤ 31000
Tare	≤ 100% of Max		
Temperature range	+10 °C / +30 °C		

Type	HJ-21 KCE & HJ-31 KCE & LN21001.. & LN31001	
Accuracy class	(II)	
Max	21 kg ~ 31 kg	100000 ~ 150000 ct
e ≥	1 g	5 ct
d ≥	0,1 g	d = e
n	≤ 31000	≤ 30000
Tare	≤ 100% of Max	
Temperature range	+10°C / +30°C	

In the description number T7345 revision 1 further characteristics are described.

Valid until 21 January 2018

Description and documentation The instrument is described in the description number T7345 revision 1 and documented in the documentation folder T7345-2, appertaining to this EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier version, including its documentation folder.

The Notified Body no. 0122
 NMI Certin, 17 June 2011



C. Oosterman
 Head Certification Board