

EC type-approval certificate

Number **T7518** revision 1 Project number 11200401 Page 1 of 4

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Giropes S.L. Pol. Emporda Intern. C/F, Parcela 15-16 17469 Vilamalla (Girona) Spain
In respect of	A class (III), electronic, single- or multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer mark/name: Giropes Type : PTI & PTI C
Characteristics	$ \begin{array}{l} 6 \ kg \leq Max \leq 30 \ kg \\ e \geq 1 \ g \\ n \leq 3000 \ divisions (per partial weighing range) \\ Maximum \ of 2 \ partial \ weighing \ ranges \\ Temperature \ range \ 0 \ ^C \ / \ +40 \ ^C \\ In \ the \ description \ number \ T7518 \ revision \ 1 \ further \ characteristics \ are \ described. \end{array} $
	24 December 2018 The instrument is described in the description number T7518 revision 1 and documented in the documentation folder T7518-1, appertaining to this EC type-approval certificate.
Remarks	This revision EC type-approval certificate replaces the earlier version, except for its documentation folder.
The Notified Bo NMi Certin, 20	
C. Oosterman	1 + + + + + + + + + + + + + + + + + + +
C. Oosterman Head Cortificat	<pre></pre>