

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant Metlab Metroloji Ltd. Sti.
Alinteri Bulvari Gul 86
Toplu Isyerleri No. 1/16
06370 Ostim – Ankara
Turkey

In respect of A class **(III)**, electronic, single- or multi-interval, **non-automatic weighing instrument**, intended to be used for direct sales to the public.
Manufacturer : Metlab
Type : QSP / QTP / QHW

Characteristics $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
 $e \geq 1 \text{ g}$
 $n \leq 3000$ divisions (per partial weighing range)
Maximum of 2 partial weighing ranges
 $T \leq -\text{Max}$
Temperature range $-10^\circ\text{C} / +40^\circ\text{C}$

In the description number T7557 revision 0 further characteristics are described.

Valid until 2 April 2019

Description and documentation The instrument is described in the description number T7557 revision 0 and documented in the documentation folder T7557-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122
NMI Certin, 2 April 2009


C. Oosterman
Head Certification Board