

## EC type-approval certificate

Number **T7707** revision 1 Project number 11200720 Page 1 of 5

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 2009/23/EC on non-automatic weighing instruments.

with

Manufacturer Dibal S.A.

Astinze Kalea, 24-Pol. Ind. Neinver 48160 Derio (Bilbao-Vizcaya)

Spain

In respect of A class (III), electronic, single- or multi-interval non-automatic weighing

**instrument**, intended to be used for direct sales to the public.

Manufacturer mark/name: Dibal, Cely

Type : WP / WSP / WPP (direct sales to the public)

SW / SWP (not for direct sales to the public)

Characteristics  $3 \text{ kg} \leq \text{Max} \leq 45 \text{ kg}$ 

e ≥ 1 g

 $n \le 3000$  divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

 $T \leq -Max$ 

Temperature range -10 °C / +40 °C

In the description number T7707 revision 1 further characteristics are described.

Valid until 15 March 2021

Description and The instrument is described in the description number T7707 revision 1 and documentation documented in the documentation folder T7707-2, appertaining to this

EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier version, including its

documentation folder.

The Notified Body no. 0122 NMi Certin, 20 December 2011

C. Øosterrhan

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl

www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted

