



# EC type-approval certificate

Number **T7780** revision 0  
Project number 10200468  
Page 1 of 6

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer Ishida Co. Ltd.  
44 Sanno-cho, Shogoin  
Sakyo-ku, Kyoto-city, 606-8392  
Japan

In respect of A class **III**, electronic, single- or multi-interval **non-automatic weighing instrument**.  
Manufacturer mark/name: Ishida  
Type : IZ-7000

### Characteristics

Weighing unit type: S...	Weighing unit type: IW...	Weighing unit type: M...
1.2 kg ≤ Max ≤ 120 kg e ≥ 0.2 g n ≤ 6000 divisions	3 kg ≤ Max ≤ 150 kg e ≥ 1 g n ≤ 3000 divisions (per partial weighing range Maximum of 2 partial weighing ranges	60 kg ≤ Max ≤ 300 kg e ≥ 20 g n ≤ 3000 divisions

T ≤ -(Max – e) for single-interval instruments

T ≤ -(Max<sub>1</sub> – e<sub>1</sub>) for multi-interval instruments

Temperature range 0 °C / +35 °C

In the description number T7780 revision 0 further characteristics are described.

Valid until 20 July 2020

Description and documentation The instrument is described in the description number T7780 revision 0 and documented in the documentation folder T7780-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122  
NMI Certin, 20 July 2010

  
C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document only is permitted

