

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer Precisa Gravimetrics AG  
Moosmattstrasse 32  
CH-8953 Dietikon  
Switzerland

In respect of A class **(II)**, electronic, single-interval **non-automatic weighing instrument**.  
Manufacturer mark/name: Precisa  
Type : BJ165 Series

Characteristics  $500\text{ g} \leq \text{Max} \leq 6000\text{ g}$   
 $0.1\text{ g} \leq e < 1\text{ g}$   
 $d = 1/10\text{ e}$  or  $d = e$   
 $n \leq 6100$  divisions  
Temperature range  $+10\text{ }^\circ\text{C} / +30\text{ }^\circ\text{C}$

In the description number T7943 revision 0 further characteristics are described.

Valid until 17 June 2021

Description and documentation The instrument is described in the description number T7943 revision 0 and documented in the documentation folder T7943-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122  
NMI Certin, 17 June 2011



C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document only is permitted