

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer GRAM PRECISION SL
Travesía Industrial 11
08907 L'Hospitalet de Llobregat
Barcelona
Spain

In respect of A class **II**, electronic, single-interval, **non-automatic weighing instrument**,
Manufacturer : Gram Precision
Type : SVI-500, SV-500ix, VXI-500, VX-500i, S1-500i

Characteristics Max \leq 510 g
 $e \geq$ 10 mg
 $d = 0.1 e$
 $T \leq -Max$
 $n \leq 51.000$ divisions
Temperature range +15 °C / +30 °C

In the description number T7976 revision 0 further characteristics are described.

Valid until 20 September 2021

Description and documentation The instrument is described in the description number T7976 revision 0 and documented in the documentation folder T7976-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122
NMI Certin, 20 September 2011



C. Oosterman
Head Certification Board