

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Excell Precision Co., Ltd. 6F, No. 127 Lane 235, Pao-Chiao Rd. Hsin-Tien, Taipei Hsien Taiwan
In respect of	A class (III) , electronic, single- or multi-interval, multi-range non-automatic weighing instrument . Manufacturer mark/name: Excell Precision Co., Ltd. Type : Depending on the configuration
Characteristics	$1,5 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$ $e \geq 0,5 \text{ g}$ $n \leq$ the number of scale intervals mentioned in the test certificate involved. Maximum of 2 partial weighing ranges Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ In the description number T8045 revision 0 further characteristics are described.
Valid until	17 November 2021
Description and documentation	The instrument is described in the description number T8045 revision 0 and documented in the documentation folder T8045-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122
NMi Certin, 17 November 2011



C. Oosterman
Head Certification Board