

| | |
|-------------------------------|--|
| Issued by | NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands |
| In accordance with | The Council Directive 2009/23/EC on non-automatic weighing instruments. |
| Manufacturer | DIGI France S.A. Z.A. Central Parc 4 Allée du Sanglier 93421 Villepinte France |
| In respect of | A class (III) , electronic, single- or multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer mark/name: DIGI France S.A. Type : SM-800., SM-880.. |
| Characteristics | $6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$ $30 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$ (instruments with external weighing platform) $e \geq 1 \text{ g}$ $e \geq 5 \text{ g}$ (instruments with external weighing platform) $n \leq 3000$ divisions (per partial weighing range) Maximum of 2 partial weighing ranges Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ In the description number T8204 revision 0 further characteristics are described. |
| Valid until | 10 July 2022 |
| Description and documentation | The instrument is described in the description number T8204 revision 0 and documented in the documentation folder T8204-1, appertaining to this EC type-approval certificate. |

The Notified Body no. 0122
NMi Certin, 10 July 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted